

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



Members of the
UK Weighing Federation
Membership No. 1070

CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

<p style="font-size: 1.2em; font-family: cursive;">Wasteology Greenham TA21 0JU</p>
LOCATION

Machine Details

Cust ref	
Make	Bilanciai
Model	D410
Ser No	J20404
Capacity	5000kg /
e/d	20kg /
Working cap	

Test Weights

Set No.	Cert No.	Date Cal
PTG 1-24	TR11126	2/12/20

Cert Interval	12 months
Next Cert	March 2022

Tolerance +/-	kg
0-10000	20
10000-40000	40
40000-50000	60

<p>Comments</p> <p style="font-size: 1.2em; font-family: cursive;">make up weight used FLT = 7520 kg Unit = 20 kg</p>	<table border="1" style="margin: auto;"> <tr><td> </td><td>X</td><td> </td></tr> <tr><td>2</td><td>4</td><td>6</td></tr> <tr><td>1</td><td>3</td><td>5</td></tr> </table>		X		2	4	6	1	3	5
	X									
2	4	6								
1	3	5								

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
0 kg	0	0	0	0
6000	6000	0	6000	0
10000	10000	0	10000	0
16000	16000	0	16000	0
20000	20000	0	20000	0
25520	25520	0	25520	0
34140	34140	0	34140	0
44140	44120	20		

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

Repeatability

Load	Ind	Ind	Ind
44140 kg	44140	44120	44120

Repeatability

Load	Ind	Ind	Ind

Eccentric Load

Reading	Dev	Reading	Dev
1 7520	0	2 7520	0
3 7520	0	4 7520	0
5 7540	+20	6 7540	+20
7		8	

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Signature Name A. Manfield Date 15/3/21